



RETAIL LEVEL RECALL

11/3/17

Dear Member,

The purpose of this letter is to advise you that Cardinal Health is voluntarily recalling a specific production lot of **ARGYLE SUCTION TUBING**.

As of July 29, 2017 Cardinal Health acquired this product line from Medtronic, and is initiating this recall of the Covidien Product below.

This recall is being conducted due to incorrect packaging. In one lot of Argyle Suction Tubing, the pouch has the potential to be incorrectly labeled as the Salem Sump Dual Lumen Stomach Tube. This product may have been distributed within the case of Argyle Suction Tubing. The use of this product may cause a delay in patient treatment while the proper tubing is retrieved.

There have been no reports from customers related to this issue and the recall affects only the products listed.

PRODUCT RECALLED BY CARDINAL HEALTH/COVIDIEN

<u>MUTUAL ITEM #</u>	<u>PRODUCT</u>	<u>ITEM CODE</u>	<u>LOT NUMBER</u>
222-869	Argyle Suction Tubing	88883015115	710016864X
N/A	Salem Sump Dual Lumen Stomach Tube	8888266148	710016864X


With this recall, you are asked to:

- Check your product for the affected lots by referring to the attachment.
- Stop dispensing and quarantine all impacted product.
- **RETURN THIS PRODUCT TO MUTUAL MARKED RECALL**
- Contact Brent Slaughter if you have additional questions.

Attachment A

Item code

Item code



Lot number:
710016864X

Lot number:
710016864X

Case level packaging label

Lot number:
710016864X



Attachment B

Correct Pouch

Incorrect Pouch

