



Nuclear Medicine Technologist Scope of Practice and Performance Standards

**Prepared by: Society of Nuclear Medicine and
Molecular Imaging Technologist Section
Approved: September 2016**

1 **Overview of Document**

2
3 This document includes the Scope of Practice and the Performance Standards for health care
4 professionals that, for the purposes of this document, will be referred to as a nuclear
5 medicine technologist.

6
7 The spectrum of responsibilities for a nuclear medicine technologist varies widely across
8 the United States. Practice components presented in this document include what is taught in
9 Nuclear Medicine programs, tested by accrediting organizations, and practiced in the field.
10 This document provides a basis for establishing the areas of knowledge and performance for
11 the nuclear medicine technologist.

12
13 The nuclear medicine technologist **MUST BE IN COMPLIANCE WITH ALL FEDERAL,**
14 **STATE, AND INSTITUTIONAL GUIDELINES** including proper documentation of initial
15 and continued competency in those practices and activities.

16
17 Continuing education is a necessary component in maintaining the skills required to perform
18 all duties and tasks of the nuclear medicine technologist in this ever-evolving field.

19
20 **Limitation of Scope and Disclaimer**

21
22 This document is intended to set forth the standards in important areas of the nuclear
23 medicine technologist’s responsibilities. It may not cover all areas which may present
24 themselves in actual practice. These standards do not supersede the judgment of the
25 individual nuclear medicine technologist and other healthcare professionals serving the
26 patient in light of all of the facts of the individual case. **THE SOCIETY OF NUCLEAR**
27 **MEDICINE AND MOLECULAR IMAGING AND THE SOCIETY OF NUCLEAR**
28 **MEDICINE AND MOLECULAR IMAGING TECHNOLOGIST SECTION DISCLAIM**
29 **ALL LIABILITY ARISING FROM USE OF THESE DOCUMENTS.**

30
31 **Overview**

32
33 Nuclear medicine is a medical technology that utilizes sealed and unsealed radioactive
34 materials for diagnostic, treatment, and research purposes. Nuclear medicine instrumentation
35 may be combined with, computed tomography (CT), magnetic resonance imaging (MRI), or
36 other modalities to produce three-dimensional images with or without adjunctive and other
37 imaging medications to enhance the evaluation of physiological processes at a molecular
38 level.

39
40 **Technologist Qualified to Perform Nuclear Medicine Procedures**

41
42 Under the supervision of an authorized user, the nuclear medicine technologist is
43 responsible for the safe use of ionizing and nonionizing radiation and molecular imaging for
44 diagnostic, therapeutic, and research purposes. The technologist will review the patient’s
45 medical history to understand the patient’s illness, medical issue, and pending diagnostic or
46 treatment procedure; instruct the patient before, during, and following the procedure;

47 evaluate the satisfactory preparation of the patient before beginning a procedure; and
48 recognize emergency patient conditions and initiate lifesaving first aid when appropriate.

49
50 Administrative functions may include supervising other technologists, students, and other
51 personnel; participating in procuring supplies and equipment; documenting laboratory
52 operations; participating in radiation safety protocols and taking an active role in radiation
53 reduction programs; participating in departmental inspections conducted by various licensing,
54 regulatory, and accrediting agencies; participating in departmental quality assurance or
55 quality improvement projects; and participating in scheduling patient procedures.

56
57 A certified nuclear medicine technologist is an individual who is registered or certified by the
58 Nuclear Medicine Technology Certification Board (NMTCB), the American Registry of
59 Radiologic Technologists (ARRT), Canadian Association of Medical Radiation
60 Technologists (CAMRT), and/or any other certification board accepted by your state or
61 institution. A certified nuclear medicine technologist is qualified to perform general nuclear
62 medicine procedures, nuclear medicine therapy, nuclear cardiology procedures, nuclear
63 breast procedures, positron emission tomography (PET) procedures, and CT attenuation
64 correction and localization at entry level. An advanced certification in CT through the
65 NMTCB, ARRT, CAMRT, and/or any other certification board accepted by your state or
66 institution qualifies a certified nuclear medicine technologist to perform diagnostic CT.

67
68 **Education**

69 Nuclear Medicine Technologists may complete a one- or two- year certificate program, a
70 two-year associate's degree, bachelor's degree or Master's Degree. Didactic courses include
71 but are not limited to the physical sciences, biological effects of radiation exposure, radiation
72 protection, radiation procedures, CT anatomy and physics, the use of radiopharmaceuticals,
73 adjunctive medications, imaging medication, imaging techniques, and computer applications.
74 A structured clinical education component provides experience in the clinical environment.
75 Clinical education is designed to meet the requirements of the certification exams. Graduates
76 of accredited programs are eligible to sit for certification examinations offered by the
77 NMTCB, ARRT, and/or any other certification board accepted by your state or institution.
78 The Joint Review Committee on Education Programs in Nuclear Medicine Technology
79 accredits training programs in nuclear medicine technology.

80
81 **Licensure**

82 Requirements for licensure of all imaging technologists vary from state to state, so it is
83 important that technologists check the requirements of the state in which they plan to work.

84
85 **Certification**

86 Certification is available from the NMTCB, ARRT, and/or any other certification board
87 accepted by your state or institution

88
89 **Continuing Education**

90 In addition to the general certification requirements, certified technologists also must
91 complete a certain number of continuing education hours to maintain certification.

92 Continuing education is required because of the frequent technological and

93 radiopharmaceutical innovations.

94

95

96

Code of Ethics

97

98 Technologists qualified to perform nuclear medicine procedures are members of the health
99 care profession and must strive as individuals and as a group to maintain the highest ethical
100 standards by adhering to the *Nuclear Medicine Technologist Code of Ethics* approved by the
101 *Society of Nuclear Medicine and Molecular Imaging Technologist Section (SNMMITS)*.

102

103 The principles of the *Nuclear Medicine Technologist Code of Ethics* as listed below are not
104 laws, but standards of conduct to be used as ethical guidelines by nuclear medicine
105 technologists.

106

Principle 1

107 The nuclear medicine technologist will provide services with compassion and respect for
108 the dignity of the individual and with the intent to provide the highest quality of patient
109 care.

110

Principle 2

111 The nuclear medicine technologist will provide care without discrimination regarding the
112 nature of the illness or disease, gender, race, religion, sexual preference, or
113 socioeconomic status of the patient.

114

Principle 3

115 The nuclear medicine technologist will maintain strict patient confidentiality in
116 accordance with state and federal regulations.

117

Principle 4

118 The nuclear medicine technologist will comply with the laws, regulations, and policies
119 governing the practice of nuclear medicine.

120

Principle 5

121 The nuclear medicine technologist will continually strive to improve his or her
122 knowledge and technical skills.

123

Principle 6

124 The nuclear medicine technologist will not engage in fraud, deception, or criminal
125 activities.

126

Principle 7

127 The nuclear medicine technologist will be an advocate for his or her profession.

128

129

130

131

132

Definitions

133 **Adjunctive Medication:** Adjunctive medications are defined as those medications used

139 to evoke a specific physiological or biochemical response used in conjunction with
140 diagnostic imaging or therapeutic procedures.

141

142 **ALARA:** ALARA is an acronym for "as low as (is) reasonably achievable," which
143 means making every reasonable effort to maintain exposures to ionizing radiation as far
144 below the dose limits as practical. *The NRC definition under 10 CFR Part 20.1003 of*
145 *ALARA can be found here:* [http://www.nrc.gov/reading-rm/basic-](http://www.nrc.gov/reading-rm/basic-ref/glossary/alara.html)
146 [ref/glossary/alara.html](http://www.nrc.gov/reading-rm/basic-ref/glossary/alara.html).

147

148 **Authorized User:** A physician licensed to permit the medical use of byproduct
149 material. *The NRC definition under 10 CFR Part 35.2 of an Authorized User can be*
150 *found here:* [//www.nrc.gov/reading-rm/doc-collections/cfr/part /part - .html](http://www.nrc.gov/reading-rm/doc-collections/cfr/part/part-.html)

151

152 **Computed Tomography:** A medical imaging technology that uses a computer to
153 acquire a volume of x-ray-based images, generally reconstructed as two-dimensional
154 (2D) or three- dimensional (3D) pictures of inside the body.

155

156 **Diagnostic Imaging:** Diagnostic imaging uses technologies such as x-ray, CT, MR,
157 ultrasound, general nuclear medicine, PET, and single-photon emission computed
158 tomography (SPECT) to provide physicians with a way to look inside the body without
159 surgery.

160

161 **Diagnostic Nuclear Medicine:** The use of radioactive materials (called
162 radiopharmaceuticals or radiotracers) to evaluate molecular, metabolic, physiologic,
163 anatomic, and pathologic conditions of the body for the purposes of diagnosis and
164 research.

165

166 **Hybrid Imaging:** The combination of imaging technologies that allows information
167 from different modalities to be presented as a single set of images.

168

169 **Imaging Device:** A technological apparatus used to produce detailed images of the
170 inside of the body for diagnostic, therapeutic, or research purposes. Examples of these
171 devices include the gamma camera, CT scanner, PET scanner, MR unit, optical imaging
172 detector, and ultrasound device.

173

174 **Imaging Medication:** Medication that is administered immediately before or
175 during an imaging procedure and is used only to enhance imaging studies. It
176 includes but is not limited to iodinated contrast and gadolinium.

177

178 **Isotope:** Atoms of a single element that have differing masses. Isotopes are either
179 stable or unstable (radioisotope). Radioisotopes are radioactive: they emit
180 particulate (alpha, beta) or electromagnetic (gamma) radiation as they transform or
181 decay into stable isotopes.

182

183 **Magnetic Resonance Imaging:** Magnetic resonance (MR) imaging is a diagnostic scan
184 that uses high-strength magnetic fields and radio frequency transmission rather than

185 ionizing radiation. MR imaging techniques are used primarily to study anatomy, but a
186 special type of MR scan, functional MR imaging (fMRI), can be used to map blood flow
187 for functional studies.

188
189 **Molecular Imaging:** Molecular imaging is an array of non-invasive, diagnostic imaging
190 technologies that can create images of physical, functional, and anatomical aspects of
191 the living body at a molecular level. Molecular imaging technologies include, but are not
192 limited to, nuclear medicine, optical imaging, spectroscopy, PET, and SPECT.

193
194 **Nuclear Medicine Therapy:** The use of radioactive materials (called
195 radiopharmaceuticals or radiotracers) to treat disease processes.

196
197 **Positron Emission Tomography:** Positron emission tomography is a medical imaging
198 technology using radiopharmaceuticals emitting positrons that annihilate into two
199 photons. These photon pairs are detected by the PET scanner to produce images.

200
201 **Radiopharmaceuticals:** Radioactive chemicals used to diagnose, treat, or prevent disease.
202

203 **Single Photon Computed Tomography:** SPECT imaging uses a gamma camera to
204 acquire multiple 2-D images (projections) from multiple angles. Tomographic
205 reconstruction algorithms are applied to the multiple projections, yielding a 3-D dataset.
206 This dataset may then be manipulated to show thin slices along any chosen axis of the
207 body, similar to those obtained from other tomographic techniques, such as CT, PET and
208 MRI.

209

210 **The Scope of Practice**

211

212 The scope of practice in nuclear medicine technology includes, *but is not*
213 *limited to*, the following areas and responsibilities:

214

215 **Patient Care:** Requires the exercise of judgment to assess and respond to the patient's
216 needs before, during, and following diagnostic imaging and treatment procedures and in
217 patient medication reconciliation. This includes record keeping in accordance with the
218 Health Insurance Portability and Accountability Act (HIPAA).

219

220 **Instrumentation/Quality Control:**

221 Involves the operation of:

222

223 Nuclear medicine and PET imaging systems:

224 With or without sealed sources of radioactive materials, x-ray tubes, or MR
225 systems for attenuation correction, transmission imaging, or diagnostic CT or
226 MR (when appropriately trained and/or credentialed).

227

228 Non-imaging

229 instrumentation:

230 Dose calibrators

231 Survey instrumentation for exposure and contamination

232 Probe and well instrumentation
233 Ancillary patient care equipment as authorized by institutional policies
234 Infusion systems
235 Radionuclide generators

236
237 Quality control:
238 The evaluation and maintenance of a quality control program for all
239 instrumentation to ensure optimal performance and stability.

240
241 **Diagnostic Procedures:** Requires the utilization of appropriate techniques,
242 radiopharmaceuticals, imaging medications and adjunctive medications as part of a
243 standard protocol to ensure quality diagnostic images and/or laboratory results.
244 Obtains biological samples to perform testing as required for the optimization of
245 patient care and quality of diagnostic procedures.

246
247 **Therapeutic Procedures:** Requires the utilization of appropriate techniques,
248 radiopharmaceuticals, and adjunctive medications as part of a standard protocol to ensure
249 proper treatment of the disease process. Obtains biological samples to perform testing as
250 required for the optimization of patient care.

251
252 **Adjunctive Medications:** Involves the identification, preparation, calculation,
253 documentation, administration, and monitoring of adjunctive medication(s) used during
254 diagnostic imaging, or therapeutic procedures.

255
256 **Imaging Medications:** Involves the identification, preparation, calculation, documentation,
257 administration, and monitoring of imaging medication(s) used during diagnostic imaging
258 studies.

259
260 **Radiopharmaceuticals:** Involves the safe handling and storage of
261 radiopharmaceuticals. This includes, but is not limited to, the procurement,
262 identification, preparation, dose calculation, and administration of
263 radiopharmaceuticals. It also includes all associated documentation and disposal as
264 appropriate.

265
266 **Radiation Safety:** Involves practicing techniques that will minimize radiation exposure
267 to the patient, health care personnel, and general public. These include using protective
268 devices, shields, dose reduction, and monitors consistent with ALARA principles.
269 Establishing protocols for managing spills and unplanned releases of radiation.

270
271

272 The Clinical Performance Standards

273

274 The clinical performance standards for the nuclear medicine technologist include,
275 *but are not limited to*, the following areas and responsibilities:

276

277 I. Patient Care

278 A. A nuclear medicine technologist prepares the patient by:

- 279 1. Verifying patient identification, date of last menstrual period, pregnancy
280 or breastfeeding status (and alerting the authorized user if there are
281 concerns about possible pregnancy), and written orders for the procedure.
282 2. Assuring study appropriateness based on indication and patient symptoms.
283 Consulting with the authorized user and/or referring physician whenever the request
284 is called into question.
285 3. Obtaining a pertinent medical history, including medications and allergies,
286 and confirming the patient’s candidacy for the procedure.
287 4. Ensuring that any pre-procedural preparation has been completed (e.g.,
288 fasting, diet, hydration, glucose levels, voiding, bowel cleansing, and
289 suspension of interfering medications).
290 5. Ensuring that informed consent has been obtained and witnessed, as prescribed
291 by the institution, whenever necessary.
292 6. Properly explaining the procedure to the patient and/or family and, where
293 appropriate, to the parent and/or legal guardian, and when necessary, obtaining
294 the assistance of an interpreter or translator. This includes, but is not limited to,
295 patient involvement, length of study, radiation safety issues, and post-
296 procedure instructions.
297
- 298 B. A nuclear medicine technologist provides patient care by:
- 299 1. Assuring comfort and care to the patient prior to, during, and following a procedure.
300 This includes, but is not limited to, the use and monitoring of intravenous lines (i.e.,
301 central lines, peripherally inserted central catheters (PICC)), oxygen supplies, and
302 drains. This also includes the operation of blood pressure cuffs, electrocardiogram
303 (ECG) machines, pulse oximeters, glucometers, intravenous pumps, and oxygen
304 delivery regulators as authorized by institutional policies.
305 2. Inserting and monitoring peripheral intravenous catheters.
306 3. Nuclear Medicine Technologists administer radioactive, adjunctive, and imaging
307 medications. This includes, but is not limited to, the following: oral, intravenous,
308 intramuscular, intradermal, subcutaneous, inhalation.
309 4. Monitoring patients who are under minimal sedation in accordance with the
310 American Society of Anesthesiologists [ASA] guidelines for conscious sedation and
311 per institutional guidelines and documenting during the monitoring period.
312 5. Collecting specimens and performing pertinent laboratory procedures. Performing in
313 vitro diagnostic testing laboratory analyses as required by established protocols.
314 Additionally, performing in vitro diagnostic testing laboratory procedures to measure
315 the biodistribution of radiopharmaceuticals.
316 6. Establishing and maintaining proper communication with patients (i.e., proper
317 introduction, appropriate explanation of procedure, etc.).
318 7. Maintaining a professional demeanor at all times to assure the preservation of
319 patients’ rights, resulting in the provision of the highest-quality patient care possible.
320 8. Following recognized infection control practices to provide a safe and sanitary
321 working environment for patients and the general public.
322 9. Recognizing and responding to an emergency situation at a level commensurate
323 with one’s training and competency, including cardiopulmonary resuscitation
324 (CPR); the use of automatic external defibrillators (AED), if applicable; advanced

325 cardiac life support (ACLS); and advanced pediatric life support (PALS).
326 10. Recognizing, responding to, reporting, and documenting adverse events.

327

328 C. A nuclear medicine technologist performs administrative procedures by:

- 329 1. Maintaining an adequate volume of medical/surgical supplies, imaging
330 medications, adjunctive medications, radiopharmaceuticals, storage media, and
331 other items required to perform procedures in a timely manner.
- 332 2. Scheduling patient procedures appropriate to the indication and in the proper
333 sequence.
- 334 3. Maintaining appropriate records of administered radioactivity, quality control
335 procedures, patient reports, and other required records.
- 336 4. Developing and revising, when necessary, policies and procedures in accordance
337 with applicable regulations.
- 338 5. Actively participating in total quality management/continuous quality
339 improvement programs (i.e., age-specific competencies, patient education, and
340 patient restraint and immobilization).
- 341 6. Complying with licensing standards and institutional policies. The nuclear
342 medicine technologist involved with research must also follow Institutional
343 Research Board protocols, comply with Institutional Animal Care and Use
344 Committee, and Food and Drug Administration standards.

345

346 **II. Instrumentation/Quality Control**

347 A. A nuclear medicine technologist evaluates equipment performance, initiates corrective
348 action when necessary, and maintains required records for the quality control program of
349 gamma camera imaging systems, PET systems, hybrid imaging systems, CT, and/or MR
350 in accordance with applicable regulations, accrediting agencies, and recommendations
351 from camera manufacturers. Responsibilities include but are not limited to:

- 352 1. Identifying system-specific quality control requirements by following
353 recommended initial acceptance quality control procedures and daily, weekly,
354 monthly, quarterly, and annual quality control procedures to evaluate allowable
355 parameter ranges for uniformity, photon detection/discrimination, spatial
356 resolution, scatter correction, count loss, measurement of random interactions,
357 sensitivity, dead-time loss, and random count correction accuracy as
358 recommended by the manufacturer, and required by institutional and
359 accreditation policies.
- 360 2. Recognizing image artifacts requiring imaging system correction and performing
361 corrections and quality assurance.
- 362 3. Performing and evaluating sinogram acquisition or other routine quality control
363 procedures to evaluate detector integrity.
- 364 4. Performing imaging system quality assurance is based upon recommendations
365 from the physicist, service engineer, and/or camera manufacturer. It includes,
366 but is not limited to:
 - 367 a. Obtaining uniformity images on imaging detectors.
 - 368 i. Selecting a radionuclide source of appropriate type, size,
369 quantity, and energy.
 - 370 ii. Selecting an appropriate pulse height analyzer (PHA), photopeak,

- 371 and window.
- 372 iii. Obtaining uniformity images using standardized imaging
- 373 parameters.
- 374 iv. Evaluating the images qualitatively and/or
- 375 quantitatively in comparison to the manufacturer's
- 376 specifications and the performance requirements based
- 377 on the studies for which the unit is used.
- 378 v. Identifying the source of any significant nonuniformity
- 379 (e.g., checking collimator and PHA peak setting).
- 380 vi. Initiating corrective action when necessary.
- 381 b. Performing a detector linearity evaluation on imaging detectors.
- 382 i. Selecting a radionuclide, selecting a linearity phantom,
- 383 and obtaining images.
- 384 ii. Identifying any nonlinear distortion in the
- 385 image.
- 386 iii. Determining the source of nonlinearity (e.g., detector-
- 387 source geometry).
- 388 iv. Initiating corrective action when necessary.
- 389 c. Performing spatial resolution checks on imaging detectors.
- 390 i. Selecting an appropriate radionuclide.
- 391 ii. Choosing a phantom that is compatible with the
- 392 specified resolution of the camera.
- 393 iii. Analyzing the resulting images for degradation of resolution
- 394 and determining the causes.
- 395 iv. Initiating corrective action when necessary.
- 396 d. Conducting sensitivity checks on imaging detectors yearly in
- 397 conjunction with a physicist.
- 398 i. Selecting a source with an appropriate level of activity and half-
- 399 life.
- 400 ii. Ensuring identical geometry, source placement, and
- 401 measurement parameters for repetitive checks.
- 402 iii. Evaluating results.
- 403 iv. Initiating corrective action when necessary.
- 404 e. Performing single-photon emission computed tomography (SPECT) quality
- 405 control procedures based on camera manufacturer recommendations,
- 406 including but not limited to:
- 407 i. Obtaining a high-count uniformity calibration flood.
- 408 ii. Obtaining a center-of-rotation calibration to ensure
- 409 detector alignment.
- 410 iii. Evaluating reconstruction results of an acquired cylindrical SPECT
- 411 phantom with contrast and spatial resolution inserts:
- 412 a. Detector quality control may include but is not limited to
- 413 the evaluation of system uniformity, sensitivity, linearity
- 414 and spatial resolution.
- 415 b. Record and evaluate results according to manufacturer
- 416 guidelines' institutional and accreditation policy.

- 417 c. Initiating corrective action when necessary.
- 418 f. Performing CT system quality assurance based on camera manufacturer
- 419 recommendations, including but not limited to:
- 420 i. Daily: Follow camera manufacturers' described warm-up procedure
- 421 and automatic monitoring, at various tube voltage (kVp) or current
- 422 (mAs) settings, of the tube output and detector response.
- 423 ii. Follow camera manufacturers' recommendations: Perform a phantom
- 424 evaluation to determine tomographic uniformity accuracy of the CT
- 425 number for water, image noise, and slice thickness.
- 426 iii. Initiating corrective action when necessary.
- 427 g. Performing PET or PET/CT system quality assurance based on camera
- 428 manufacturer recommendations, including but not limited to:
- 429 i. Acquiring consistent 2D and/or 3D PET images, using appropriate
- 430 reconstruction techniques, to display sinogram images for QC
- 431 interpretation.
- 432 ii. Working in conjunction with medical director or medical
- 433 physicists verifying CT/AC protocols, including mAs, kVp, pitch,
- 434 and helical scanning.
- 435 iii. Initiating corrective action when necessary.
- 436 5. Performing radionuclide generator quality assurance, daily and before the use of the
- 437 generator, to include dose calibrator/generator calibration and parent/daughter
- 438 breakthrough.
- 439 6. Performing infusion device quality control per manufacturer recommendations.
- 440 7. Operating imaging systems, storage media, and radiation detection and counting
- 441 devices, including but not limited to imaging detectors, dose calibrators, survey
- 442 instruments, scintillation probes, well counters, and data processing and image
- 443 production devices:
- 444 a. Maintaining and operating auxiliary equipment used in procedures.
- 445 b. Actively participating in total quality management/continuous quality
- 446 improvement programs by:
- 447 i. Identifying indicators to be analyzed.
- 448 ii. Gathering and presenting data in appropriate formats, analyzing
- 449 data, and recommending changes.
- 450 8. Operating scintillation probes, well counters, and other laboratory equipment:
- 451 a. Calibrating a spectrometer with a long-half-life radionuclide source.
- 452 b. Determining energy resolution.
- 453 c. Conducting sensitivity and constancy measurements at appropriate
- 454 energies with a standard, long-lived source Cs-137 or I-129.
- 455 d. Checking background and determining the cause for levels greater than
- 456 established normal levels.
- 457 e. Conducting a chi-square test.
- 458 f. Maintaining required records for quality control programs in
- 459 accordance with federal and state regulations and institutional policies.
- 460 g. Performing glucometer quality assurance using high and low standards.
- 461 9. Operating survey meters:
- 462 a. Ensuring that calibration has been completed within the last 12 months.

- 463 b. Performing a battery check to verify the meter is operational.
 464 c. Performing a check-source test and comparing with previous results.
 465 d. Maintaining required records for the quality control program.
 466 10. Operating dose calibrator:
 467 a. Verifying constancy every day that isotopes are administered to patients,
 468 including weekends and on-call hours, and checking channels of the
 469 isotopes used that day using a check source with a long half-life.
 470 b. Verifying linearity quarterly over the entire range of radionuclide activity
 471 to be administered to patients, comparing calculated activities to measured
 472 activities, and determining correction factors when necessary.
 473 Determining accuracy annually by comparing a set of known activities to
 474 measured activities using isotopes of varying energy emissions such as
 475 Co-57, Ba-133, and Cs-137.
 476 c. Upon installation, testing for significant geometric variation in activity
 477 measured as a function of sample volume or configuration and
 478 determining correction factors when necessary.
 479 d. Maintaining required records for the quality control program in
 480 accordance with federal and state regulations and institutional policies.
 481 11. Operating image processors/computer monitors:
 482 a. Verifying the calibration of the instrument.
 483 b. Maintaining required records for the quality control program.
 484

485 III. Diagnostic Procedures

- 486 A. A nuclear medicine technologist performs imaging procedures by:
 487 1. Determining appropriate imaging parameters.
 488 a. Preparing (see Section V.C.), evaluating, and properly administering the
 489 prescribed amount of various radiopharmaceuticals, adjunctive
 490 medications, and/or imaging medications.
 491 b. Selecting the appropriate imaging or data collection parameters.
 492 2. Administering radiopharmaceuticals, adjunctive medications, and/or imaging
 493 medications through various routes (including but not limited to oral, intravenous,
 494 intramuscular, intradermal, subcutaneous, inhalation) in accordance with
 495 established protocols and verifying that the radiopharmaceutical meets quality
 496 specifications prior to administration (i.e., expiry time, pH, half-life, etc.).
 497 3. Administering radiopharmaceuticals, adjunctive medications, and imaging
 498 medications:
 499 a. Verifying patient ID according to institutional policy.
 500 b. Determining route of administration according to established protocol.
 501 c. Establishing and/or verifying venipuncture access using aseptic technique.
 502 d. Using and maintaining established venous access routes (e.g., heparin
 503 infusion set, infusion pump, peripherally inserted central catheter (PICC),
 504 and central line).
 505 e. Reconciling patient medications according to institutional policy to ensure
 506 that the patient's current medications will not interact with the
 507 radiopharmaceutical, adjunctive medications, and imaging medications
 508 used for the ordered exam.

- 509 f. Preparing (see Section IV.C.) and administering adjunctive medications
510 and imaging medications per the appropriate route.
- 511 g. Documenting medications and/or radiopharmaceutical administrations in
512 the patient medical record in accordance with federal and state regulations
513 and institutional policies.
- 514 h. Observing the patient carefully after any administration for side effects,
515 and handling such side effects appropriately as described in established
516 policies or as directed by medical staff.
- 517 4. Positioning the patient and obtaining images:
- 518 a. Verifying energy peak on NM cameras.
- 519 b. Waiting an appropriate time following the administration of a
520 radiopharmaceutical, adjunctive medication, or imaging medication to
521 begin the imaging procedure protocol, and acquiring additional views as
522 necessary to optimize information content.
- 523 c. Exercising professional judgment in positioning a patient to best
524 demonstrate pathology and to adapt to the patient's limitations.
- 525 d. Positioning the patient using supportive materials and immobilizers, as
526 necessary.
- 527 e. Indicating appropriate anatomic landmarks for each view of the
528 procedure.
- 529 f. Reviewing images to ensure that the required information has been
530 acquired and that the images have been processed properly and are of
531 the highest quality.
- 532 5. Assisting in exercise and pharmacologic cardiac testing procedures:
- 533 a. Preparing patients to include the correct placement of ECG electrodes.
- 534 b. Determining if the appropriate test has been ordered based on the ECG
535 rhythm, medical history, and current medications.
- 536 c. Recognizing and responding to ECG changes.
- 537 d. Recognizing the parameters that indicate termination of a cardiac stress
538 study.
- 539 e. Recognizing ECG patterns that are appropriate for image gating.
- 540 6. Performing data collection, processing, and analysis:
- 541 a. Performing data collection, processing, and analysis in accordance with
542 institutional protocols.
- 543 b. Exercising independent judgment in selecting appropriate images for
544 processing.
- 545 c. Obtaining quantitative measurements such as SUV, coronary flow reserve,
546 kinetic modeling, regional brain analysis, biliary and cardiac ejection
547 fractions, and renal function, as appropriate for the procedure performed.
- 548 d. Defining regions of interest (ROIs) with reproducible results and correctly
549 applying background subtraction.
- 550 e. Performing computer data manipulations as required.
- 551 f. Labeling processed images (e.g., anatomical positioning, ROIs, date, and
552 time).
- 553 g. Archiving to and retrieving data from storage media.
- 554

- 555 B. A nuclear medicine technologist may perform non-imaging in vitro and/or radioassay
556 studies by:
- 557 1. Operating laboratory equipment, including well counters, probes, and other
558 detection devices to measure the biodistribution of radiopharmaceuticals.
 - 559 2. Preparing doses:
 - 560 a. Quantitating doses:
 - 561 i. Calculating and confirming the activity to be used
 - 562 ii. Calculating the volume necessary to deliver activity for the
563 prescribed dose.
 - 564 iii. Preparing standard solutions or dosage for phantom use as
565 needed using appropriate volumetric or gravimetric
566 techniques to dilute the standard per institutional protocol.
 - 567 3. Collecting appropriate biological specimens for procedures using standard
568 precaution techniques as required by protocol:
 - 569 a. Collecting blood samples:
 - 570 i. Selecting proper supplies (e.g., needles, syringes, evacuated tubes,
571 or anticoagulants).
 - 572 ii. Identifying and verifying the patient and labeling patient
573 demographics on collection containers.
 - 574 iii. Performing venipuncture at appropriate intervals using aseptic
575 technique.
 - 576 iv. Adding hemolyzing compounds or anticoagulants to samples
577 according to protocol.
 - 578 v. Centrifuging blood and separating blood components, according to
579 protocol.
 - 580 vi. Storing aliquots of serum, plasma, or whole blood according to
581 protocol.
 - 582 b. Collecting urine samples by:
 - 583 i. Instructing the patient and/or nursing staff regarding the correct
584 method and time of urine collection.
 - 585 ii. Aliquoting the urine sample and measuring total urine volume.
 - 586 iii. Measuring the specific gravity of urine, if required.
 - 587 iv. Recognizing and documenting all technical circumstances that
588 would produce invalid results
 - 589 4. Gathering, validating, and documenting data:
 - 590 a. Subtracting room background or patient background from appropriate
591 samples.
 - 592 b. Applying appropriate formulas, including conversion and dilution factors.
 - 593 c. Calculating results according to the procedure used.
 - 594 d. Plotting a graph, if necessary, and determining half time by extrapolating
595 to zero time.
 - 596 e. Reporting both calculated values for a patient and normal range of specific
597 procedures used.
 - 598 f. Evaluating results for potential error.
 - 599 5. Managing biohazardous, chemical, and radioactive waste in accordance with
600 applicable state and federal regulations and institutional policy.

601

602 **IV. Adjunctive Medications**

603 A nuclear medicine technologist displays:

604 A. A thorough understanding and knowledge of indications, contraindications, warnings,
605 precautions, proper use, drug interactions, and adverse reactions for each adjunct
606 medication to be used.

607

608 B. The ability to procure and maintain adjunctive medications and supplies by:

609 1. Anticipating and procuring a sufficient supply of medications for an appropriate
610 period in accordance with anticipated need.

611 2. Storing medications and supplies in a manner consistent with labeled product
612 safeguards and established institutional policies.

613 3. Identifying and properly disposing of expired medications.
614

615 C. The ability to properly prepare and administer adjunctive medications under the
616 supervision of an authorized user by:

617 1. Employing aseptic technique for manipulation of sterile products and
618 preparations.

619 2. Selecting and preparing adjunctive medications.

620 3. Confirming the quality of an adjunctive medication in accordance with accepted
621 techniques and official standards.

622 4. Documenting the administered dose, date, and time of all adjunctive medications
623 in a permanent medical record.

624 5. Observing the patient for possible complications (e.g., adverse reactions) of
625 adjunctive medication administration, and handling such complications
626 appropriately in conjunction with other available staff.
627

628 **V. Imaging Medications**

629 A nuclear medicine technologist displays:

630 A. A thorough understanding and knowledge of indications, contraindications, warnings,
631 precautions, proper use, drug interactions, and adverse reactions for each imaging
632 medication to be used.
633

634 B. The ability to procure and maintain imaging medications and supplies by:

635 1. Anticipating and procuring a sufficient supply of medications for an appropriate
636 period in accordance with anticipated need.

637 2. Storing medications and supplies in a manner consistent with labeled product
638 safeguards and established institutional policies.

639 3. Identifying and properly disposing of expired medications.
640

641 C. The ability to properly prepare and administer imaging medications under the
642 supervision of an authorized user by:

643 1. Employing aseptic technique for manipulation of sterile products and
644 preparations.

645 2. Selecting and preparing imaging medications in accordance with the
646 manufacturer's specifications and institutional policy.

647 3. Confirming the quality of an imaging medication in accordance with accepted

- 648 techniques and official standards.
 649 4. Documenting the administered dose, date, and time of all imaging medications in
 650 a permanent medical record.
 651 5. Observing the patient for possible complications (e.g., adverse reactions) of
 652 imaging medication administration, and handling such complications
 653 appropriately in conjunction with other available staff.
 654

655 **VI. Radiopharmaceuticals**

656 A. A nuclear medicine technologist displays a:

- 657 1. Thorough knowledge of indications, contraindications, warnings, precautions,
 658 proper use, drug interactions, and adverse reactions for each radiopharmaceutical
 659 to be used.
 660 2. Thorough knowledge of biochemical and molecular functions that relate to, but
 661 not limited to, glucose metabolism, blood flow, brain oxygen utilization,
 662 perfusion, and receptor–ligand binding rates.
 663 3. Thorough knowledge of the physiological and biochemical processes that
 664 relate to organ system function and anatomy and radiopharmaceutical
 665 demonstration of normal and pathologic states.
 666

667 B. A nuclear medicine technologist maintains radiopharmaceutical products by:

- 668 1. Anticipating and procuring a sufficient supply of radiopharmaceuticals for an
 669 appropriate period in accordance with anticipated need and license possession
 670 limits.
 671 2. Maintaining security while storing radiopharmaceuticals in a manner consistent
 672 with the manufacturer’s labeled product safeguards, radiation safety
 673 considerations, and established policies.
 674 3. Performing and documenting radiation survey and wipe tests upon receipt of
 675 radioactive materials.
 676 4. Recording receipt of radioactive materials in a permanent record.
 677 5. Following Department of Transportation (DOT) regulations and radiation safety
 678 guidelines in the transport, receipt, and shipment of radioactivity.
 679

680 C. A nuclear medicine technologist properly prepares and administers

681 radiopharmaceuticals under the direction of an authorized user in accordance with all
 682 federal and state regulations and institutional policies by:

- 683 1. Preparing all sterile radiopharmaceuticals in appropriate environments in compliance
 684 with USP and FDA Standards.
 685 2. Following appropriate personnel cleansing and garbing protocols when entering
 686 “clean” areas in accordance with USP Standards.
 687 3. Employing aseptic technique, consistent with USP Standards, when mixing and
 688 manipulating sterile products
 689 4. Following appropriate USP Standards for beyond-use date (time-of-use) and vial
 690 puncture standards.
 691 5. Assembling and maintaining radionuclide generators.
 692 6. Eluting radionuclide generators according to the manufacturer’s specification in a
 693 “clean” environment that complies with USP Standards.

- 694 7. Verifying the radionuclidic purity of generator eluates.
- 695 8. Selecting and preparing radiopharmaceuticals in accordance with the
- 696 manufacturer's specifications.
- 697 9. Measuring the radioactivity of the radiopharmaceutical using a dose calibrator.
- 698 10. Confirming the quality of a radiopharmaceutical in accordance with accepted
- 699 techniques and official standards (e.g., radiochemical purity and physical
- 700 appearance).
- 701 11. Handling and preparing blood or blood products for labeling and/or labeled blood
- 702 cells in accordance with established regulations and protocols and in an
- 703 environment in compliance with USP Standards, and ensuring that when blood
- 704 products are handled and compounded they are separated from other
- 705 radiopharmaceuticals.
- 706 12. Recording use and/or disposition of all radioactive materials in a permanent
- 707 record:
- 708 a. Properly storing radiopharmaceutical kits, and radiopharmaceuticals as
- 709 stated in USP Standards.
- 710 b. Recording results of radionuclide generator eluates' quality assurance tests
- 711 to include dose calibrator/generator calibration and radionuclidic purity of
- 712 eluates.

713
714 D. A nuclear medicine technologist is responsible for the identification and labeling of all
715 radiopharmaceutical preparations by:

- 716 1. Labeling vials and syringes.
- 717 2. Recording radiopharmaceutical and medication information on a patient's
- 718 administration form and permanent preparation records.
- 719 3. Labeling and segregating radioactive waste and recording the information in a
- 720 permanent record.

721
722 E. A nuclear medicine technologist prepares individual dosages under the supervision of
723 an authorized user by:

- 724 1. Applying radioactive decay calculations to determine the required volume or unit
- 725 form necessary to deliver the prescribed radioactive dose.
- 726 2. Selecting and preparing prescribed dosages and entering the information on a
- 727 patient's administration form and other permanent records.
- 728 3. Appropriately labeling the dose for administration.
- 729 4. Checking the dose activity prior to administration in a dose calibrator and
- 730 comparing this measurement against the shipment documentation.
- 731 5. Recording use and/or disposition of radioactive materials in a permanent
- 732 record by properly storing radiopharmaceuticals.

733
734 **VII. Radionuclide Therapy**

735 A. A nuclear medicine technologist properly prepares and/or administers therapeutic
736 radiopharmaceuticals when these agents are part of a standard procedure that is required
737 for treatment under the direct supervision of an authorized user by:

- 738 1. Ensuring that the correct radiopharmaceutical and dosage is prepared.
- 739 2. Following the quality management program in effect at the facility in regard to

- 740 patient identification and verification and the use of therapeutic
741 radiopharmaceuticals.
- 742 3. Observing prescribed radiation safety using FDA and USP Standards during the
743 preparation and administration of such treatment.
- 744 4. Assisting the authorized user in supplying proper patient care instructions to
745 hospital staff, patient, and/or caregivers.
- 746 5. Conducting and documenting radiation surveys of designated patient areas, when
747 indicated.
- 748 6. Instructing the patient, family, and staff in radiation safety precautions after the
749 administration of therapeutic radiopharmaceuticals.
- 750 7. Coordinating/scheduling pre-/post treatment blood/urine draws and/or imaging.
- 751 8. Maintaining all appropriate records.

752
753 **VIII. Radiation Safety**

754 A. A nuclear medicine technologist performs all procedures utilizing ionizing radiation
755 safely and effectively by:

- 756 1. Maintaining security of radioactive materials.
- 757 2. Notifying the appropriate authority when changes occur in the radiation safety
758 program.
- 759 3. Assisting in the preparation of license amendments when necessary.
- 760 4. Keeping up to date on regulatory changes and complying with all applicable
761 regulations.
- 762 5. Maintaining required records.
- 763 6. Posting appropriate radiation signage in designated areas.
- 764 7. Following federal and state regulations regarding receipt, storage, disposal, and
765 usage of all radioactive materials.
- 766 8. Recommending the purchase of radiation protection equipment to meet federal
767 and state regulations and institutional policies.
- 768 9. Packaging and monitoring radioactive material for transport according to federal
769 and state regulations, and keeping accurate records of transfer.

770
771 B. A nuclear medicine technologist follows appropriate radiation protection procedures
772 by:

- 773 1. Using personnel monitoring devices (film badges, optically stimulated
774 luminescence [OSL] thermoluminescent dosimeters, etc.).
- 775 a. Reviewing personnel exposure records in regard to maximum
776 permissible dose limits.
- 777 b. Taking appropriate measures to reduce exposure.
- 778 c. Notifying proper authorities of excessive exposure
779 upon discovery/occurrence.
- 780 2. Selecting and using proper syringe shields and other shielding configurations to
781 reduce radiation exposure to patients, personnel, and the general public.
- 782 3. Using proper shielding and disposal procedures to maximize patient, technologist,
783 and public protection.
- 784 4. Working in a safe but timely manner in order to decrease radiation exposure in
785 consideration of ALARA guidelines.

- 786 5. Reviewing personnel monitoring device readings to determine if radiation
787 exposure can be further reduced.
- 788 6. Working in a manner that minimizes potential contamination of patients,
789 technologists, the public, and work areas.
- 790
- 791 C. A nuclear medicine technologist monitors for radioactive contamination at
792 regular intervals or after repairs by:
- 793 1. Ensuring that instruments are calibrated.
- 794 2. Setting the frequency and locations for surveys and following schedules.
- 795 3. Using appropriate survey meters for each type and level of activity.
- 796 4. Following federal and state regulations regarding personnel surveys and reporting
797 to the designated authorized user or radiation safety officer.
- 798 5. Performing constancy checks on survey meters.
- 799 6. Performing wipe tests where applicable.
- 800 7. Performing leak tests on sealed sources.
- 801 8. Recording data in the required format (e.g., dpm instead of cpm).
- 802 9. Evaluating the results of wipe tests and area surveys to determine if action is
803 required.
- 804 10. Notifying the radiation safety officer when actions are required.
- 805
- 806 D. A nuclear medicine technologist performs decontamination procedures by:
- 807 1. Wearing personal protective equipment as necessary.
- 808 2. Restricting access to the affected area and confining a spill.
- 809 3. Removing contamination and monitoring the area and personnel, and repeating
810 the decontamination procedure until activity levels are acceptable.
- 811 4. Closing off all areas of fixed contamination that are above acceptable levels,
812 shielding the area, and posting appropriate signs.
- 813 5. Identifying, storing, or disposing of contaminated material.
- 814 6. Maintaining appropriate decontamination records.
- 815 7. Notifying the appropriate authority (e.g., radiation safety officer) in the event of
816 possible overexposure or other violations of federal and state regulations and
817 institutional policies.
- 818
- 819 E. A nuclear medicine technologist disposes of radioactive waste by:
- 820 1. Maintaining appropriate records.
- 821 2. Disposing according to license specifications.
- 822 3. Maintaining radioactive storage areas.
- 823 4. Maintaining current Hazmat training records per NRC and Organization of
824 Agreement States (OAS) regulations.
- 825 F. A nuclear medicine technologist participates in programs designed to instruct other
826 personnel about radiation hazards and principles of radiation safety by:
- 827 1. Using the following teaching concepts:
- 828 a. Types of ionizing radiation.
- 829 b. Biological effects of ionizing radiation.
- 830 c. Limits of dose, exposure, and radiation effect.

- 831 d. Concepts of low-level radiation and health.
832 e. Concept of risk versus benefit.
833 f. ALARA
- 834 2. Providing appropriate radiation safety measure instructions.
835 3. Providing proper emergency procedures instruction.
836 4. Modeling proper radiation safety techniques and shielding in the course of daily
837 duties.
- 838
- 839 G. A nuclear medicine technologist assists in performing radiation safety procedures
840 associated with radionuclide therapy by:
- 841 1. Following the guidelines for administration of therapeutic radiopharmaceuticals
842 and the release of patients administered therapeutic radiopharmaceuticals.
843 2. Following the guidelines for the release of patients administered radioactive
844 materials.
845 3. Following the proper procedures for patients requiring hospitalization after
846 administration of therapeutic radiopharmaceuticals.
847 4. Providing appropriate instruction on radiation safety procedures for patients, care
848 givers, and staff.
849

850

851

852

References

853

854 Accreditation Standards for Nuclear Medicine Technologist Education. 2011.

855 [://www.jrcnmt.org/essentials.asp](http://www.jrcnmt.org/essentials.asp).

856

857 American College of Radiology. Technical Guidelines for Nuclear Medicine and
858 Radiopharmaceuticals. 2006.

859 [://www.acr.org/SecondaryMainMenuCategories/quality_safety/ guidelines/tech-](http://www.acr.org/SecondaryMainMenuCategories/quality_safety/guidelines/tech-standardsnm.aspx)

860 [standardsnm.aspx](http://www.acr.org/SecondaryMainMenuCategories/quality_safety/guidelines/tech-standardsnm.aspx).

861

862 American Registry of Radiologic Technologists. Content Specifications.

863 [://www.arrt.org/examinations/contentspecs/NMT_CS_2011.pdf](http://www.arrt.org/examinations/contentspecs/NMT_CS_2011.pdf)

864

865 American Registry of Radiologic Technologists. Task Inventory for Nuclear Medicine
866 Technologist. [://www.arrt.org/pdfs/Examinations/NMT-Task-Inventory.pdf](http://www.arrt.org/pdfs/Examinations/NMT-Task-Inventory.pdf). Published July
867 2010. Last accessed January 19, 2012.

868

869 American Society of Radiologic Technologists. Nuclear Medicine Practice Standards. 2010

870 [://www.asrt.org/media/pdf/practicestds/GR10_OPI_Strds_NM_PS.pdf](http://www.asrt.org/media/pdf/practicestds/GR10_OPI_Strds_NM_PS.pdf)

871

872 American Society of Radiologic Technologists. Nuclear Medicine Practice Standards.

873 [://www.asrt.org/media/pdf/standards_nm.pdf](http://www.asrt.org/media/pdf/standards_nm.pdf). Last accessed January 19, 2012.

874

875 Blondeau K, Harten S, Pickett M, Bridges J. Critical Task Analysis Report. Tucker,
876 Georgia: Nuclear Medicine Technology Certification Board; 2000.

877

878 Bureau of Labor Statistics. Occupational Outlook Handbook, 2010-11 Edition. Nuclear
879 Medicine

880 Technologists. [://www.bls.gov/oco/ocos104.htm](http://www.bls.gov/oco/ocos104.htm)

881

882 Glossary of Molecular Imaging Terms [://interactive.snm.org/index.cfm?PageID=11120](http://interactive.snm.org/index.cfm?PageID=11120)

883

884 Joint Review Committee on Educational Programs in Nuclear Medicine
885 Technology. Accreditation Standards for Nuclear Medicine Technologist
886 Education.

887 [://jrcnmt.org/sites/jrcnmt/uploads/documents/Accred](http://jrcnmt.org/sites/jrcnmt/uploads/documents/Accred)

888 [_Policy_Documents/Standards10_2011.pdf](http://jrcnmt.org/sites/jrcnmt/uploads/documents/Accred_Policy_Documents/Standards10_2011.pdf) f. Published 2010. Updated May 2011. Last
889 accessed January 19, 2012.

890

891 Nuclear Medicine Technology Certification Board. Components of Preparedness. Nuclear
892 Medicine Exam: Components of Preparedness. [://www.nmtcb.org/exam/cops.php](http://www.nmtcb.org/exam/cops.php). Updated
893 2009. Last accessed January 19, 2012.

894

895 Nuclear Medicine Technology Certification Board (NMTCB) Report: Equipment and

- 896 Procedures in Current Practice (2002).
897
- 898 Performance and Responsibility Guidelines for the Nuclear Medicine Technologist,
899 Socio- Economic Affairs Committee, SNMITS (September 1994).
900
- 901 Performance and Responsibility Guidelines for the Nuclear Medicine Technologist,
902 Socio- Economic Affairs Committee, SNMITS (September 1998).
903
- 904 SNMITS Socioeconomic Affairs Committee. Performance and Responsibility Guidelines for
905 the
906 Nuclear Medicine Technologist. *J Nucl Med Technol.* 2003;31:222–229.
907
- 908 SNMITS Scope of Practice Task Force. Nuclear Medicine Technologist Scope of Practice
909 (September 2008).
910
- 911 SNMITS Educators Task Force Curriculum Subcommittee. NMT Entry-Level Curriculum
912 Guide,
913 4th Edition (August 2008).
914
- 915 SNMITS Position Paper. Baccalaureate Degree proposed as entry level educational
916 requirements.
917 2005. [://interactive.snm.org/index.cfm?PageID=4715](http://interactive.snm.org/index.cfm?PageID=4715)
- 918 Performance and Responsibility Guidelines for the Nuclear Medicine Technologist,
919 Socio- Economic Affairs Committee, SNMITS (September).
920
- 921 SNMITS Socioeconomic Affairs Committee. Performance and Responsibility Guidelines for
922 the
923 Nuclear Medicine Technologist. *J Nucl Med Technol.* ; : – .
924
- 925 SNMITS Scope of Practice Task Force. Nuclear Medicine Technologist Scope of Practice
926 (September).
927
- 928 SNMITS Educators Task Force Curriculum Subcommittee. NMT Entry-Level Curriculum
929 Guide,
930 th Edition (August).
931
- 932 SNMITS Position Paper. Baccalaureate Degree proposed as entry level educational
933 requirements.
934 . [://interactive.snm.org/index.cfm?PageID=](http://interactive.snm.org/index.cfm?PageID=)