

DICKSON

930 South Westwood Avenue - Addison, IL 60101
Phone: 630-543-3747 - Fax: 630-543-0498

Certificate of Validation

The undersigned ensures that the software / product system:

TM725

Hardware Product

SAMPLE

Serial Number(s)

DicksonWare

Software Product

13

Version Number

was developed, tested and validated and was found to meet the functional specifications for this product, using the procedures in effect at the time of shipment. In addition, the hardware supplied with the software, as specified, has been tested against NIST traceable standards according to established procedures to ensure that its performance meets specification.

To support this certification for GLP/GMP requirements of this product, Dickson will make the following documents available to authorized governmental/regulatory agencies for inspection at Dickson, Addison, Illinois:

- Functional Specification
- Source Code Documentation
- Operational Qualification Procedures
- Operational Qualification Results
- Software/Hardware Revision History
- D100/D200 Calibration Procedure
- Logger Calibration Procedures - Product family specific
- N525: Dickson IQOQ VALidation Package
- 21 CFR Part 11

Dickson will maintain possession of all documents and their reproductions and may require a non-disclosure agreement to be provided by those requiring access to these documents.

01/01/2008

Date

Steve Thompson

Calibration/Production Coordinator

Dan Gawel

Quality/Manufacturing Engineer