

BEFORE USING THE DEVICE, PLEASE READ THE FOLLOWING INSTRUCTIONS COMPLETELY AND CAREFULLY. CORRECT APPLICATION IS VITAL TO THE PROPER FUNCTIONING OF THE DEVICE.

WARNINGS AND CAUTIONS:

Before using this device, refer to the Contraindications, Warnings and Cautions on the side of this package. If you experience any unusual reactions while using this product, discontinue use and consult your medical professional immediately.

INTENDED USER PROFILE: THE USER SHOULD BE ABLE TO READ AND UNDERSTAND THE DIRECTIONS, WARNINGS AND CAUTIONS.

INTENDED USE/INDICATIONS:

Recommended to help relieve symptoms of:

- Chronic venous insufficiency
- Heavy, achy, fatigued legs
- Swelling in the legs, ankles and feet
- Varicose and spider veins
- Venous leg ulcers

Wearing compression stockings may be beneficial for travelers who are unable to move around for extended periods of time, for pregnant women during and after pregnancy, and for someone who is recovering from surgery.

APPLICATION INFORMATION:

Consult your health care provider for recommended frequency and duration of compression stocking application. Put on stockings first thing in the morning. Your legs have the least amount of swelling early in the morning. Remove stockings before going to bed.

CLEANING INSTRUCTIONS:



1. Wash in warm water, by hand or in gentle machine cycle, using a delicate detergent. Do not bleach or use fabric softener.



2. Machine dry on low heat, or hang to dry. Do not wring product.



3. Never iron or dry clean.

CONTRAINDICATIONS

Compression wear should not be worn if you've been diagnosed with any of these conditions:

- Advanced arterial disease
- Congestive heart failure
- Inflammation due to bacterial infection
- Severe deep venous thrombosis

WARNINGS AND CAUTIONS

Consult your healthcare professional about compression wear if you commonly experience:

- Skin infections
- Weeping dermatoses
- Neuropathy or insensitivity
- Physical immobility



Gently pull stocking over your toes, instep and heel. Do not bunch fabric together.



Gradually work straight up over your calf. The top band of Knee length stockings should sit 1-2 inches below the knee bend.



Smooth out all wrinkles and folds.

NOTICE: WHILE EVERY EFFORT HAS BEEN MADE IN STATE-OF-THE-ART TECHNIQUES TO OBTAIN MAXIMUM COMPATIBILITY OF FUNCTION, STRENGTH, DURABILITY AND COMFORT, THERE IS NO GUARANTEE THAT INJURY WILL BE PREVENTED THROUGH THE USE OF THIS PRODUCT.

WARRANTY: DJO, LLC warrants to the purchaser that any Dr. Comfort® products purchased will be free from defects in materials and workmanship for a period of 6 months from the date of sale. For complete details, visit: drcomfort.com. To the extent the terms of this warranty are inconsistent with local regulations, the provisions of such local regulations will apply.