


Wrist Brace


APPLICATION INSTRUCTIONS

CP020420 Rev A 04/08


This device is offered for sale by or on the order of a physician or other qualified medical professional. This device is not intended for reuse on a second patient. This device is not intended to prevent injuries or reduce or eliminate the risk of injury.




1. Position the Brace Place the hand into the brace with the thumb inside the thumb hole. The palmer stay should be resting on the palmer section of the hand.



2. Secure Straps Beginning at the hand and working up the arm, feed each strap through the corresponding D-ring and secure.



4. Trim Straps If necessary trim straps. The brace should be snug but comfortable.





INDICATIONS: Indicated for Carpal Tunnel Syndrome and/or immobilization of the wrist.

CONTRAINDICATIONS: Contraindicated for unstable fractures of the wrist.

WARNINGS: Do not leave this device in the trunk of a car on a hot day or in any other place where the temperature may exceed 150°F (65°C) for any extended period of time. If any additional pain or symptoms occur while using this brace, discontinue use and seek medical attention.

WARRANTY: This device is warranted to be free from defects in material and workmanship for a period of 90 days. This warranty applies to devices that have not been modified or subjected to misuse, abuse or neglect. This device is expressly subject to Bledsoe's Product Line Limited Warranty. Copies of this limited warranty can be obtained from your medical professional, via the internet at www.BledsoeBrace.com, or by calling 1.888.BLEDSOE (253.3763).

MANUFACTURED BY: BLEDSOE BRACE SYSTEMS™

2601 Pinewood; Grand Prairie, Texas, 75051, USA Toll Free Tel: 1.888.BLEDSOE (253.3763) • 1.800.527.3666 Local: 972.647.0884 • Local Fax: 972.660.5495

International Fax: 972.606.0649 www.BledsoeBrace.com

EUROPEAN AUTHORIZED REPRESENATIVE



EMERGO EUROPE Molenstraat 15 2513 BH, The Hague The Netherlands Tel: (+31) 70 345 8570 Fax: (+31) 70 346 7299