

NL6000 BASIC HANDPIECE MAINTENANCE



Remove bur from the NL6000.



Wash the NL6000 under warm water and clean softly with a toothbrush.

Manual

OR


Care3 Plus™

Insert the PANA spray tip into the drive air hole of the NL6000 and spray for 2 - 3 seconds or until lubrication is expelled from the head of the handpiece.



Open the door and properly connect the NL6000 to the standard port and make certain the bur hole is facing the door. Close the door securely.

Purge the handpiece of excess lubrication either via an air station or by returning to the operatory and placing the handpiece back on the tubing. Purging excess lubricant for up to 10 seconds is sufficient.



Select "S" mode for the short cleaning cycle. Press the START button and wait until Care3 Plus completes cleaning, lubrication and purging cycle.



PERIODIC MAINTENANCE

Push-button chuck: Approximately once a week it is beneficial to spray the handpiece chuck for 3-5 seconds to insure optimum performance. Depress push button while spraying with PANA spray using pointed nozzle.



Bag and place handpiece in the autoclave. (Remember: Do not autoclave with bur in the chuck)

It is not necessary to spray the handpiece after the autoclave cycle is complete.

Handpiece questions?
Contact Brasseler USA directly!
1.800.841.4522 (ext. 4675)

By Your Side in Dentistry



BRASSLER USA
DENTAL INSTRUMENTATION



To order call 800.841.4522
or fax 888.610.1937.
Visit our website: BrasselerUSA.com

