

Manual Supplement

Manual Title: 9007 Users
Part Number:
Print Date:
Revision/Date: 582202

Supplement Issue: **1**
Issue Date: 6/14
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 10, prior to the **Warranty** section, add:

Note

The Product can show some control sensitivity to moderate or severe electromagnetic fields or conducted interference of certain frequencies. In the presence of radiated EM disturbances, with frequencies of 250 MHz to 400 MHz and with amplitude >1 V/m to a maximum of 3 V/m, add 0.0025 °C to the stability specification. Stability is not guaranteed if amplitude is >3 V/m. When subject to conducted disturbances of 8 MHz to 80 MHz, and amplitude >3 V, add 0.005 °C to the stability specification.