



VERIFICATION OF COMPLIANCE

This is to certify that the product listed below was (were) tested in the BTL EMC Laboratory to comply with the required criteria levels of the follow-mentioned Generic Standards or Product Family Standard(s) and/or Basic Standard(s) based-on the essential conformity requirements of MDR Regulation (EU) 2017/745.

Equipment Industrial PC
Model Name XXXXXCL260XXXXXXXXXXXXXXX (where 'X' can be any alphanumeric character, "-" or blank designating configuration differences)
Brand Name ONLOGIC
Applicant OnLogic, Inc.
Address 435 Community Dr South Burlington, VT 05403-6860 United States
Standard(s) EN 60601-1-2:2015+A1:2021
BS EN 60601-1-2:2015+A1:2021
IEC 60601-1-2:2014+AMD1:2020 Ed. 4.1
(for the professional healthcare facility environment)
EN 55011:2016+A2:2021 Class A Group 1
BS EN 55011:2016+A2:2021 Class A Group 1
CISPR 11:2015+AMD1:2016+AMD2:2019 Ed. 6.2
EN IEC 61000-3-2:2019+A2:2024 Class A
BS EN IEC 61000-3-2:2019+A2:2024 Class A
EN 61000-3-3:2013+A2:2021+AC:2022-01
BS EN 61000-3-3:2013+A2:2021+AC:2022-01
Report(s) BTL-EMC-2-2510T051
Issued Date 2025/12/3

The test data, data evaluation, and equipment configuration contained in our test report(s) above was (were) obtained utilizing the test procedures, test instruments, test sites that has been accredited by the Authority of TAF according to the ISO/IEC 17025 quality assessment standard and technical standard(s). The test data contained in the referenced test report relate only to the EUT sample and item(s) tested.



Jack Kao

Authorized Signatory



BTL INC.

No.18, Ln. 171, Sec. 2, Jiuzong Rd.,
Neihu Dist., Taipei City 114, Taiwan
TEL:+886-2-2657-3299
FAX:+886-2-2657-3331

