

HD120™ MLC High-Definition Multileaf Collimator Bibliography*

For Stereotactic Radiosurgery (SRS) and Stereotactic Body Radiotherapy (SBRT)

Wen N, Li H, Song K, Chin-Snyder K, Qin Y, Kim J, Bellon M, Gulam M, Gardner S, Doemer A, Devpura S, Gordon J, Chetty I, Siddiqui F, Ajlouni M, Pompa R, Hammoud Z, Simoff M, Kalkanis S, Movsas B, Siddiqui MS. [Characteristics of a novel treatment system for linear accelerator-based stereotactic radiosurgery.](#) *J Appl Clin Med Phys.* 2015 Jul 8;16(4):5313. Henry Ford Health System, Detroit, MI

Subramanian SV, Subramani V, Thirumalai Swamy S, Gandhi A, Chilukuri S, Kathirvel M. [Is 5 mm MMLC suitable for VMAT-based lung SBRT? A dosimetric comparison with 2.5 mm HDMLC using RTOG-0813 treatment planning criteria for both conventional and high-dose flattening filter-free photon beams.](#) *J Appl Clin Med Phys.* 2015 Jul 8;16(4):5415. Bharathiar University, India

Serna A, Puchades V, Mata F, Ramos D, Alcaraz M. [Influence of multi-leaf collimator leaf width in radiosurgery via volumetric modulated arc therapy and 3D dynamic conformal arc therapy.](#) *Phys Med.* 2015 May;31(3):293-6. Santa Lucia University Hospital, Murcia, Spain

Bergman AM, Gete E, Duzenli C, Teke T. [Monte Carlo modeling of HD120™ multileaf collimator on Varian TrueBeam® linear accelerator for verification of 6X and 6X FFF VMAT SABR treatment plans.](#) *J Appl Clin Med Phys.* 2014 May 8;15(3):4686. BC Cancer Agency, Vancouver, British Columbia

Chae SM, Lee GW, Son SH. [The effect of multileaf collimator leaf width on the radiosurgery planning for spine lesion treatment in terms of the modulated techniques and target complexity.](#) *Radiat Oncol.* 2014 Mar 8;9(1):72. Cheju Halla General Hospital, Jeju, Korea

Pommer T, Falk M, Poulsen PR, Keall PJ, O'Brien RT, Munck af Rosenschöld P. [The impact of leaf width and plan complexity on DMLC tracking of prostate intensity modulated arc therapy.](#) *Med Phys.* 2013 Nov;40(11):111717. Rigshospitalet, Copenhagen

Tanyi JA, Doss EJ, Kato CM, Monaco DL, Zmeng L, Chen Y, Kubicky CD, Marquez CM, Fuss M. [Dynamic conformal arc cranial stereotactic radiosurgery: implications of multileaf collimator margin on dose-volume metrics.](#) *Br J Radiol.* 2012 Nov;85(1019):e1058-66. Oregon Health and Science University, Portland, OR

Kielar KN, Mok E, Hsu A, Wang L, Luxton G. [Verification of dosimetric accuracy on the TrueBeam® STx: rounded leaf effect of the high definition MLC.](#) *Med Phys.* 2012 Oct;39(10):6360-71. Stanford University Medical Center, Palo Alto, CA

Vazquez-Quino LA, Massingill B, Shi C, Gutierrez A, Esquivel C, Eng T, Papanikolaou N, Stathakis S. [Monte Carlo modeling of a Novalis Tx™ Varian 6 MV with HD-120 multileaf collimator.](#) *J Appl Clin Med Phys.* 2012 Sep 6;13(5):3960. University of Texas Health Science Center at San Antonio, San Antonio, TX

Kim J, Wen N, Jin JY, Walls N, Kim S, Li H, Ren L, Huang Y, Doemer A, Faber K, Kunkel T, Balawi A, Garbarino K, Levin K, Patel S, Ajlouni M, Miller B, Nurusev T, Huntzinger C, Schulz R, Chetty IJ, Movsas B, Ryu S. [Clinical commissioning and use of the Novalis Tx™ linear accelerator for SRS and SBRT.](#) *J Appl Clin Med Phys.* 2012 May 10;13(3):3729. Henry Ford Health System, Detroit, MI

Wang L, Kielar KN, Mok E, Hsu A, Dieterich S, Xing L. [An end-to-end examination of geometric accuracy of IGRT using a new digital accelerator equipped with onboard imaging system.](#) *Phys Med Biol.* 2012 Feb 7;57(3):757-69. Stanford University Medical Center, Palo Alto, CA

Borges C, Zarza-Moreno M, Heath E, Teixeira N, Vaz P. [Monte Carlo modeling and simulations of the High Definition \(HD120\) micro MLC and validation against measurements for a 6 MV beam.](#) *Med Phys.* 2012 Jan;39(1):415. MedicalConsult SA, Lisboa, Portugal

Fujimoto S, Ono K, Furukawa K, Kudo T, Akagi Y, Koyama T, Hirokawa Y, Kawai S, Nakashima T. [Influence of high-definition multileaf collimator for three-dimensional conformal radiotherapy and intensity-modulated radiotherapy of prostate cancer.](#) *Nihon Hoshasen Gijutsu Gakkai Zasshi.* 2012;68(7):825-34. Hiroshima Heiwa Clinic, Hiroshima, Japan

Ohtakara K, Hayashi S, Tanaka H, Hoshi H. [Dosimetric comparison of 2.5 mm vs. 3.0 mm leaf width micro-multileaf collimator-based treatment systems for intracranial stereotactic radiosurgery using dynamic conformal arcs: implications for treatment planning.](#) *Jpn J Radiol.* 2011 Nov;29(9):630-38. Gifu University Graduate School of Medicine, Gifu, Japan

* This bibliography is a comprehensive selection of articles but is not necessarily an exhaustive list of literature pertaining to SRS and SBRT.

Fix MK, Volken W, Frei D, Frauchiger D, Born EJ, Manser P. Monte Carlo implementation, validation, and characterization of a 120 leaf MLC. *Med Phys*. 2011 Oct;38(10):5311-20. Inselspital and University of Bern, Bern, Switzerland

Tanyi JA, Kato CM, Chen Y, Chen Z, Fuss M. Impact of the high-definition multileaf collimator on linear accelerator-based intracranial stereotactic radiosurgery. *Br J Radiol*. 2011 Jul;84(1003):629-38. Oregon Health and Science University, Portland, OR

Sharma DS, Dongre PM, Mhatre V, Heigrujam M. Physical and dosimetric characteristic of high-definition multileaf collimator (HDMLC) for SRS and IMRT. *J Appl Clin Med Phys*. 2011 Apr 14;12(3):3475. Kokilaben Dhirubhai Ambani Hospital, Mumbai, India

Dhabaan A, Elder E, Schreibmann E, Crocker I, Curran WJ, Oyesiku NM, Shu HK, Fox T. Dosimetric performance of the new high-definition multileaf collimator for intracranial stereotactic radiosurgery. *J Appl Clin Med Phys*. 2010 Jun 21;11(3):3040 Emory University, Atlanta, GA

Clift C, Thomas A, Adamovics J, Chang Z, Das I, Oldham M. Toward acquiring comprehensive radiosurgery field commissioning data using the PRESAGE/optical-CT 3D dosimetry system. *Phys Med Biol*. 2010 Mar 7;55(5):1279-93. Duke University Medical Center, Durham, NC

Tanyi JA, Summers PA, McCracken CL, Chen Y, Ku LC, Fuss M. Implications of a high-definition multileaf collimator (HD-MLC) on treatment planning techniques for stereotactic body radiation therapy (SBRT): a planning study. *Radiat Oncol*. 2009 Jul 10;4:22. Oregon Health & Science University, Portland, OR

Wu QJ, Wang Z, Kirkpatrick JP, Chang Z, Meyer JJ, Lu M, Huntzinger C, Yin FF. Impact of collimator leaf width and treatment technique on stereotactic radiosurgery and radiotherapy plans for intra- and extracranial lesions. *Radiat Oncol*. 2009 Jan 21;4(1):3. Duke University, Durham, NC

Wuu CS, Kessel J, Xu Y. 3-D dosimetric evaluation of 2.5 mm HD120™ multileaf system for intensity modulated stereotactic radiosurgery using optical CT based polymer gel dosimetry. *J. Phys.: Conf. Ser.* 2009 164(1):2-6. Columbia University, New York, NY

Fogliata A, Clivio A, Nicolini G, Vanetti E, Cozzi L. Intensity modulation with photons for benign intracranial tumours: a planning comparison of volumetric single arc, helical arc and fixed gantry techniques. *Radiother Oncol*. 2008 Dec;89(3):254-62. Oncology Institute of Southern Switzerland, Bellinzona, Switzerland

Chang Z, Wang Z, Wu QJ, Yan H, Bowsher J, Zhang J, Yin FF. Dosimetric characteristics of Novalis Tx™ system with high definition multileaf collimator. *Med Phys*. 2008 Oct;35(10):4460-3. Duke University Medical Center, Durham, NC

VARIAN
medical systems

A partner for **life**

© 2012, 2014, 2015 Varian Medical Systems, Inc. All rights reserved.
Varian, Varian Medical Systems, and TrueBeam are registered trademarks,
and HD120 and Novalis Tx are trademarks of Varian Medical Systems, Inc.

RAD 10226B

USA Headquarters, California

Varian Medical Systems
Palo Alto, CA
Tel: 650.424.5700
800.544.4636
Fax: 650.493.5637
varian.com

**Headquarters Europe, Eastern
Europe, Africa, Middle & Near East**

Varian Medical Systems
International AG
Cham, Switzerland
Tel: 41.41.749.8844
Fax: 41.41.749.8899
email: info.europe@varian.com

12/2015