



<<Date>>

<<First Name>> <<Last Name>>

<< Address line 1>>

<< Address line 2>>

<<City>> <<ST>> << Zip Code>>

IMPORTANT RECALL INFORMATION

Dear <<First Name>> <<Last Name>>,

We are sending you information about PosiFlush™ Syringe 10 mL saline flush syringes that may be valuable to you.

On April 16, 2020, BD issued a recall of PosiFlush Syringe 10 mL saline flush syringes. **This recall was issued because the product may have holes in the packaging which may represent a potential health hazard or safety risk to plan members using product affected by this recall.**

Our records show you may have received a prescription for this medicine recently from your retail pharmacy.

A list of the affected lot numbers and expiration dates is provided at the end of this letter.

To see if you have affected product, please check the lot number. The lot number is on manufacturer's label on the carton and on each syringe. If your product is not from one of the affected lot numbers, it is not affected by this recall. If your product is from one of these affected lot numbers, BD advises consumers to discard of any affected product and contact BD toll-free at 1-888-364-2985, Monday through Friday, 8:00 AM to 5:00 pm (ET) for instructions on how to receive replacement product at no charge.

Please see the enclosed BD consumer letter provided for more information regarding this recall.

Please call your doctor right away for advice if you might be using affected product. Your doctor knows your medical history and can tell you what to do next. If you need a prescription for a different medicine, please contact your doctor.

For more information, please call BD toll-free at 1-888-364-2985, Monday through Friday, 8:00 AM to 5:00 pm (ET) or visit www.bd.com. You may also call the United States Food and Drug Administration toll-free at 1-888-INFO-FDA (1-888-463-6332) or visit www.fda.gov.

(over)

This document contains references to brand-name prescription drugs that are trademarks or registered trademarks of pharmaceutical manufacturers not affiliated with CVS Caremark. Your privacy is important to us. Our employees are trained regarding the appropriate way to handle your private health information. This information is not a substitute for medical advice or treatment. Talk to your doctor or health care provider about this information and any health-related questions you have. CVS Caremark assumes no liability whatsoever for the information provided or for any diagnosis or treatment made as a result of this information.

- This material is available in other languages. Please call our member services department at the toll-free number listed on your benefit ID card. TTY users should call 1-800-863-5488.
- Esta información está disponible en otros idiomas. Por favor llame a nuestro departamento de servicios al cliente al número gratuito que aparece en su tarjeta de identificación. Los usuarios de equipo teleescritor (TTY) deben llamar al 1-800-863-5488.

Sincerely,

CVS Caremark® Medical Affairs

Affected Lots
8353952 exp. 12/31/21,
9011582 exp. 12/31/21,
9017875 exp. 12/31/21,
9024676 exp. 01/31/22,
9045702 exp. 01/31/22,
9060999 exp. 02/28/22,
9079716 exp. 02/28/22,
9127571 exp. 02/28/22,
9143529 exp. 04/30/22,
9156595 exp. 05/31/22,
9163601 exp. 05/31/22