



<<Date>>

<<First Name>> <<Last Name>>

<< Address line 1>>

<< Address line 2>>

<<City>> <<ST>> << Zip Code>>

### **IMPORTANT RECALL INFORMATION**

Dear <<First Name>> <<Last Name>>,

On August 27, 2021, Cardinal Health issued a recall of Monoject™ Flush Prefilled Syringes (0.9% Sodium Chloride). **This recall was issued because the product was found to have the potential to introduce air into the syringe during use.**

**These products may represent a potential health hazard or safety risk to patients administered product affected by this recall.**

Our records show you may have received a prescription the following product recently from your retail pharmacy:

[Drug Name]

A complete list of the affected products is provided at the end of this letter.

If you received one of these affected products from your participating retail pharmacy, please contact the pharmacy that filled your prescription for more information, including instructions.

**Please call your doctor right away for advice if you might be using affected product.** Your doctor knows your medical history and can tell you what to do next. If you need a prescription for a different product, please contact your doctor.

For more information, please call Cardinal Health toll-free at 1-888-444-5440. You may also call the U.S. Food and Drug Administration toll-free at 1-888-INFO-FDA (1-888-463-6332) or visit [www.fda.gov](http://www.fda.gov).

- This material is available in other languages. Please call our member services department at the toll-free number listed on your benefit ID card. TTY users should call 1-800-863-5488.
- Esta información está disponible en otros idiomas. Por favor llame a nuestro departamento de servicios al cliente al número gratuito que aparece en su tarjeta de identificación. Los usuarios de equipo teleescritor (TTY) deben llamar al 1-800-863-5488.

Sincerely,

This document contains references to brand-name prescription drugs that are trademarks or registered trademarks of pharmaceutical manufacturers not affiliated with CVS Caremark. Your privacy is important to us. Our employees are trained regarding the appropriate way to handle your private health information. This information is not a substitute for medical advice or treatment. Talk to your doctor or health care provider about this information and any health-related questions you have. CVS Caremark assumes no liability whatsoever for the information provided or for any diagnosis or treatment made as a result of this information.

CVS Caremark® Medical Affairs

<b>Affected Product</b>	<b>NDC #</b>	<b>Distribution Dates</b>
Monoject™ Flush Prefilled Syringes (0.9% Sodium Chloride), 12mL Syringe, 10 mL Saline Fill *	08881-5701-21	All lots distributed between 7/2019 and 7/2021
Monoject™ Flush Prefilled Syringes (0.9% Sodium Chloride), 12mL Syringe, 3mL Saline Fill	08881-5701-23	All lots distributed between 7/2019 and 7/2021
Monoject™ Flush Prefilled Syringes (0.9% Sodium Chloride), 12mL Syringe, 5mL Saline Fill	08881-5701-25	All lots distributed between 7/2019 and 7/2021