



4164-01-P

DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 868

[Docket No. FDA-2017-N-6568]

Medical Devices; Anesthesiology Devices; Classification of the External Negative Pressure
Airway Aid

AGENCY: Food and Drug Administration, HHS.

ACTION: Final order.

SUMMARY: The Food and Drug Administration (FDA or we) is classifying the external negative pressure airway aid into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for the external negative pressure airway aid's classification. We are taking this action because we have determined that classifying the device into class II (special controls) will provide a reasonable assurance of safety and effectiveness of the device. We believe this action will also enhance patients' access to beneficial innovative devices, in part by reducing regulatory burdens.

DATES: This order is effective [INSERT DATE OF PUBLICATION IN THE *FEDERAL REGISTER*]. The classification was applicable on December 23, 2015.

FOR FURTHER INFORMATION CONTACT: Todd Courtney, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 2530, Silver Spring, MD, 20993-0002, 301-796-6371, Todd.Courtney@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:

I. Background

Upon request, FDA has classified the external negative pressure airway aid as class II (special controls), which we have determined will provide a reasonable assurance of safety and effectiveness. In addition, we believe this action will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial distribution before May 28, 1976, is automatically classified as, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (see 21 U.S.C. 360c(f)(1)). We refer to these devices as "postamendments devices" because they were not in commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (21 U.S.C. 360c(i)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate by means of the procedures for premarket notification under section 510(k) of the FD&C Act and part 807 (21 U.S.C. 360(k) and 21 CFR part 807, respectively).

FDA may also classify a device through "De Novo" classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act. Section 207 of the Food and

Drug Administration Modernization Act of 1997 established the first procedure for De Novo classification (Pub. L. 105-115). Section 607 of the Food and Drug Administration Safety and Innovation Act modified the De Novo application process by adding a second procedure (Pub. L. 112-144). A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a 510(k) for a device that has not previously been classified. After receiving an order from FDA classifying the device into class III under section 513(f)(1) of the FD&C Act, the person then requests a classification under section 513(f)(2).

Under the second procedure, rather than first submitting a 510(k) and then a request for classification, if the person determines that there is no legally marketed device upon which to base a determination of substantial equivalence, that person requests a classification under section 513(f)(2) of the FD&C Act.

Under either procedure for De Novo classification, FDA shall classify the device by written order within 120 days. The classification will be according to the criteria under section 513(a)(1) of the FD&C Act. Although the device was automatically placed within class III, the De Novo classification is considered to be the initial classification of the device.

We believe this De Novo classification will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens. When FDA classifies a device into class I or II via the De Novo process, the device can serve as a predicate for future devices of that type, including for 510(k)s (see 21 U.S.C. 360c(f)(2)(B)(i)). As a result, other device sponsors do not have to submit a De Novo request or premarket approval application in order to market a substantially equivalent device (see 21 U.S.C. 360c(i), defining "substantial equivalence").

Instead, sponsors can use the less-burdensome 510(k) process, when necessary, to market their device.

II. De Novo Classification

On August 18, 2014, Sommetrics submitted a request for De Novo classification of the cNEP Airway Management System. FDA reviewed the request in order to classify the device under the criteria for classification set forth in section 513(a)(1) of the FD&C Act.

We classify devices into class II if general controls by themselves are insufficient to provide reasonable assurance of safety and effectiveness, but there is sufficient information to establish special controls that, in combination with the general controls, provide reasonable assurance of the safety and effectiveness of the device for its intended use (see 21 U.S.C. 360c(a)(1)(B)). After review of the information submitted in the request, we determined that the device can be classified into class II with the establishment of special controls. FDA has determined that these special controls, in addition to the general controls, will provide reasonable assurance of the safety and effectiveness of the device.

Therefore, on December 23, 2015, FDA issued an order to the requester classifying the device into class II. FDA is codifying the classification of the device by adding 21 CFR 868.5105. We have named the generic type of device external negative pressure airway aid, and it is identified as a prescription device that applies negative pressure to a patient's neck to aid in providing a patent airway during procedures requiring anesthesia.

FDA has identified the following risks to health associated specifically with this type of device and the measures required to mitigate these risks in table 1.

Table 1.--External Negative Pressure Airway Aid Risks and Mitigation Measures

Identified Risks	Mitigation Measures
Impaired blood flow	Clinical performance testing
Failure of device or negative pressure mechanism	Non-clinical performance testing
Adverse tissue reaction	Biocompatibility
Dislodging of plaque, leading to possible stroke	Labeling
Inadequate collar fit	Labeling
Use error	Labeling

FDA has determined that special controls, in combination with the general controls, address these risks to health and provide reasonable assurance of safety and effectiveness. For a device to fall within this classification, and thus avoid automatic classification in class III, it would have to comply with the special controls named in this final order. The necessary special controls appear in the regulation codified by this order. This device is subject to premarket notification requirements under section 510(k) of the FD&C Act.

At the time of classification, external negative pressure airway aids are for prescription use only. Prescription devices are exempt from the requirement for adequate directions for use for the layperson under section 502(f)(1) of the FD&C Act (21 U.S.C. 352(f)(1)) and 21 CFR 801.5, as long as the conditions of 21 CFR 801.109 are met.

III. Analysis of Environmental Impact

The Agency has determined under 21 CFR 25.34(b) that this action is of a type that does not individually or cumulatively have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

IV. Paperwork Reduction Act of 1995

This final order establishes special controls that refer to previously approved collections of information found in other FDA regulations. These collections of information are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of

1995 (44 U.S.C. 3501-3520). The collections of information in the guidance document “De Novo Classification Process (Evaluation of Automatic Class III Designation)” have been approved under OMB control number 0910-0844; the collections of information in part 814, subparts A through E, regarding premarket approval, have been approved under OMB control number 0910-0231; the collections of information in part 807, subpart E, regarding premarket notification submissions, have been approved under OMB control number 0910-0120; and the collections of information in 21 CFR part 801, regarding labeling, have been approved under OMB control number 0910-0485.

List of Subjects in 21 CFR Part 868

Medical devices.

Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 868 is amended as follows:

PART 868--ANESTHESIOLOGY DEVICES

1. The authority citation for part 868 is revised to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 360l, 371.

2. Add § 868.5105 to subpart F to read as follows:

§ 868.5105 External negative pressure airway aid.

(a) *Identification.* An external negative pressure airway aid is a prescription device that applies negative pressure to a patient’s neck to aid in providing a patent airway during procedures requiring anesthesia.

(b) *Classification.* Class II (special controls). The special controls for this device are:

(1) Clinical performance testing must document any adverse events observed during clinical use, including impaired blood flow, and demonstrate that the device performs as intended under anticipated conditions.

(2) Non-clinical performance testing must demonstrate that the device performs as intended under anticipated patient positions, does not fail during use, and does not lose negative pressure capability. The following testing should be performed:

(i) Ability of the device to maintain a seal during various patient positions;

(ii) Device leakage testing to demonstrate the device maintains vacuum;

(iii) Drop testing to ensure the device does not incur functional damage after dropping the device; and

(iv) Functional testing after high and low storage temperature.

(3) All patient contacting components must be demonstrated to be biocompatible.

(4) Labeling must include:

(i) A summary of clinical testing results, including any adverse events and evidence that effectiveness has been achieved.

(ii) Technical specifications of the device, including collar sizes, maximum duration of use, operating temperature, and storage temperature range.

(iii) Technical specifications of the vacuum source, including maximum vacuum level and operational vacuum level.

(iv) Instructions for use that includes how to place the device, determination of size, verification of suction, reference to training materials, and information on troubleshooting the device if it does not attach properly.

(v) A warning to screen patients for carotid artery disease due to the probable risk of the device to dislodge arterial plaques in the carotid artery.

(vi) A warning to exclude patients with anatomical abnormalities.

(vii) A warning not to use the device during medical procedures involving medications that contain propofol.

Dated: December 20, 2017.

Leslie Kux,

Associate Commissioner for Policy.

[FR Doc. 2017-27784 Filed: 12/22/2017 8:45 am; Publication Date: 12/26/2017]