

Yiwu EO Sterilization Technology CO., Ltd.

义乌市亿欧灭菌技术有限公司

地址：浙江省义乌市佛堂镇双峰路 517 号 Building 5, No. 517 Shuangfeng Road, Fotang Town, Yiwu city, Zhejiang

编号 No.: WT-2024013


Certification of EO Gas Sterilization

灭菌证明

Products Name 产品名称	Type/Size 型号规格	LOT No.批号	Quantity 数量
见附页	见附页	见附页	126CTNS
Sterilization agent:灭菌方法	Ethylene Oxide (100%)		
Sterile No.:灭菌批号	24051122		
Sterilization date:灭菌日期	2024.05.11		
Precondition parameters 预处理参数		Specification 操作要求	Reached values 实际值
Internal chamber temperature during action time (°C) 温度		50±10	50.5-52.9
Internal chamber Reactive humidity during action time (%)湿度		40~80%	47.7~73.2
Internal chamber time during action time (min)时间		720±30	720
Sterilization parameters 灭菌参数			
Internal chamber temperature during action time (°C) 温度		50±5	50.9~53.4
Internal chamber Reactive humidity during action time (%)湿度		40~80%	51.5~72.2
Pressure difference before and after injection EO (Kpa)压力		30±2	30.0
Agent concentration (mg/L)浓度		500±50	488.7
Duration faction time (minutes)灭菌时间		360±10	360
Flushing cycles (times)清洗次数		≥3	3
Outgassing parameters 解析参数		☒ N/A ☐ If applicable	
Internal chamber temperature during action time (°C) 温度		50±10	
Internal chamber time during action time (hr.)时间		≥24	
Evaluation of sterilization parameters 结论		☒ conformity ☐ non-conformity	
The above products are sterilized by EO gas normally, and realized sterility.以上产品由环氧乙烷灭菌, 已经灭菌完成。			

Compiled by: 

Date: 2024.05.18.

Approved by: 

Date: 2024.5.18.

附页:

Yiwu EO Sterilization Technology CO., Ltd.

义乌市亿欧灭菌技术有限公司

地址: 浙江省义乌市佛堂镇双峰路 517 号 Building 5, No. 517 Shuangfeng Road, Fotang Town, Yiwu city, Zhejiang

Goods name	Lot	ITEM	CTNS	BOXES
	012024	C0403RL	1	100
	012024	C0603RL	2	200
	012023	C1201RL	2	200
	012024	C1203RL	10	1000
	012024	C1205RL	5	500
	012024	C1207RL	5	500
	012024	C1209RL	5	500
	012024	C1218RL	1	100
	012024	C1203RLT	4	400
	012024	C1205RLT	4	400
	012024	C1003RLT	2	200
	012024	C1005RLT	1	100
	012024	C1007RLT	1	100
	012024	C1009RLT	1	100
	012024	C0809RLT		0
	012024	C1203RS-1	2	200
	012024	C1205RS-1	2	200
	012024	C1207RS-1	2	200
	012024	C1208RS-1	2	200
	012024	C1209RS-1	2	200
	012024	C1205M1-1	1	100
	012024	C1207M1-1	3	300
	012024	C1209M1-1	3	300
	012024	C1211M1-1	1	100
	012024	C1213M1-1	1	100
	012024	C1215M1-1	1	100
	012024	C1217M1-1	1	100
	012024	C1223M1-1	1	100
	012024	C1007M1-1	1	100

Yiwu EO Sterilization Technology CO., Ltd.

义乌市亿欧灭菌技术有限公司

地址: 浙江省义乌市佛堂镇双峰路 517 号 Building 5, No. 517 Shuangfeng Road, Fotang Town, Yiwu city, Zhejiang

PAPA Cartridges V1	012023	C1009M1-1	4	400
	012024	C1011M1-1	2	200
	012024	C1013M1-1	1	100
	012024	C1015M1-1	2	200
	012024	C1017M1-1	1	100
	012024	C1027M1-1	1	100
	012024	C809M1-1	1	100
	012024	C811M1-1	1	100
	012024	C813M1-1	1	100
	012024	C815M1-1	1	100
	012024	C817M1-1	1	100
	012024	C823M1-1	1	100
	012024	C827M1-1	1	100
	012024	C1205M1C-1	2	200
	012024	C1207M1C-1	4	400
	012024	C1209M1C-1	3	300
	012024	C1211M1C-1	1	100
	012024	C1213M1C-1	1	100
	012024	C1215M1C-1	1	100
	012024	C1217M1C-1	2	200
	012024	C1223M1C-1	1	100
	012024	C1007M1C-1	1	100
	012023	C1009M1C-1	4	400
	012024	C1011M1C-1	2	200
	012024	C1013M1C-1	1	100
	012024	C1015M1C-1	2	200
	012024	C1017M1C-1	1	100
	012024	C1027M1C-1	2	200

Yiwu EO Sterilization Technology CO., Ltd.

义乌市亿欧灭菌技术有限公司

地址：浙江省义乌市佛堂镇双峰路 517 号 Building 5, No. 517 Shuangfeng Road, Fotang Town, Yiwu city, Zhejiang

	012023	C807M1C-1	1	99
	012024	C809M1C-1	1	100
	012024	C811M1C-1	1	100
	012024	C813M1C-1	1	100
	012024	C815M1C-1	1	100
	012023	C817M1C-1	1	96
	012024	C823M1C-1	1	100
	012024	C827M1C-1	1	100
	012024	CP1203RL	1	100
	012024	CP1205RL	1	100
	012024	CP1207RL	1	100
	012024	CP1209RL	1	100
	012024	CH1206RL	1	100
	012024	CH1208RL	1	100
	012024	CH1210RL	1	100
Total1			128	12795

2024.05.18

