

MODEL: 353000- SINGLE PATIENT STATION



353000

FEATURES

- Multiple call-in priorities available
- Speaker and microphone provide full-duplex audio to speak with the patient
- Tilt/Release DIN receptacle for Pillow Speakers with patient entertainment controls and handset intercom with entertainment muting during nurse call
- Tilt/Release DIN receptacle allows for ease of disconnecting Pillow Speaker from the patient station
- Auxiliary device receptacle with separate call-in priority setting
- No dummy plugs required
- Interfaces with specialty beds via separate receptacle
- LED to indicate specialty bed is connected
- Interfaces with a remote Pillow Speaker via separate station
- LED to indicate remote Pillow Speaker is connected
- Output for lighting control options direct to low voltage lighting controller
- Monitor LED indicates audio status
- Cancel button for localized call cancellation
- Continuous supervision of station for loss of data and/or power faults
- Plug-in connectors for easy service
- Bio-seal option to cover button and allow for easy disinfecting of the station

SPECIFICATIONS

Power Requirements: 220mA at 18V (full load all connections)

Weight: 0.6 lbs. (0.25 kg)

Size: W: 6.4" (16.2 cm)

H: 4.25" (10.8 cm)

D: 2.0" (5.1 cm) in back box; 0.5" (1.3 cm) off-wall

Housing and Finish: High-impact molded Cyclooy 2950 UL 94-5VB; light gray faceplate

Controls/Indicators: One (1) DIN receptacles, One (1) ¼" receptacle, One (1) Cancel button, One (1) Monitor LED,

DESCRIPTION

The Single Patient Station allows both patients and staff to place different call-ins, depending on their requirements, to the assigned nurse console(s), staff terminal(s), along with other annunciation devices such as wireless phones. The patient station interfaces with entertainment handsets and specialty beds to initiate calls, control entertainment, control lighting, and provide nurse call audio and entertainment audio through the respective speakers. When used with the Feature Bed Control Module, the audio capabilities can be

Two (2) Call-assurance LED indicators, One (1) Special Bed LED indicator, and One (1) Remote Pillow Speaker LED Indicator

Terminations: Plug-on connector

Backbox Requirement: 3 gang (Non-gangable), 3-1/2" Deep

transferred to the bed rail with entertainment muting during nurse communications.

The patient station is connected to its associated corridor light or domeless room controller and is continually supervised for both power and data. Through programming, a message can annunciate at the associated console(s) and / or the remote network status application if a problem occurs.

*Architect and Engineer (A&E) Specifications available online at: customerconnection.rauland.com
 Specifications subject to change without notice*

©Copyright 2008 Rauland-Borg Corporation Printed in USA Rev 08/2008