

**MODEL: 354000 - PULL CORD STATION WITH AUDIO**



354000

**FEATURES**

- Plastic pull-cord for easy activation
- Smooth, non-porous PVC coating on the cord provides an easily cleaned surface
- Large, easy to grab bell
- Call-in button provides an additional call-in priority with built-in Call-Assurance LED
- Speaker and microphone provide full-duplex audio to the bathroom to speak with the patient if desired
- Monitor LED and Call-Assurance LED with Cancel button
- Monitor LED indicates microphone activation
- Cancel button for localized call cancellation
- Continuous supervision of station
- Plug-in connectors for easy service
- Bio-seal option to cover buttons and allow for easy disinfecting of the station

**SPECIFICATIONS**

**Power Required:** 200mA at 18V

**Size:** W: 4.7" (11.9 cm)

H: 4.25" (10.8 cm)

D: 1.7" (4.3 cm)

**Housing and Finish:** High-impact molded Cycloy 2950 UL 94-5VBABS94VO; light grey faceplate

**Controls/Indicators:** One (1) pull cord for call actuation, One (1) Push button for call actuation, Two (2) Call-assurance LED, One (1) Monitor LED, One (1) Cancel button

**Terminations:** Plug-on connector

**Backbox Requirement:** 1 gang (Non-gangable), 3-1/2" Deep. Please note station trims to 2 gang faceplate dimensions. Allow mounting trim area.

**DESCRIPTION**

The Pull Cord station with Audio is an emergency pull cord station that provides the ability to place a unique call-in priority from a single station by pulling on its plastic cord. In addition, a call can be placed from the push button on the station that can be customized to a unique call-in priority to function as a routine call or an emergency call. The Pull Cord station with Audio also contains a speaker and microphone, so that caregivers can speak directly with the patient in the bathroom if desired. The Pull-cord station is typically located in bathrooms near the toilet. The cancel button allows for cancellation of the call at the station. Unit may utilize remote Cancel button(s) for ease of nurse call cancel when pull cord is mounted in an inaccessible location (e.g. ceiling).

The Pull-cord station is connected to its associated corridor light or domeless room controller and is continually supervised.

*Architect and Engineer (A&E) Specifications available online at: [customerconnection.rauland.com](http://customerconnection.rauland.com)  
 Specifications subject to change without notice*

©Copyright 2008 Rauland-Borg Corporation Printed in USA Rev 08/2008