

MODEL: R4KSAR - SLIM PUSH FOR HELP / STAFF REGISTRATION OR CHECK-IN-STATION



R4KSAR

FEATURES

- Two (2) call points with Call-Assurance LEDs and Monitor LED
- Push for Help
- Staff Registration OR
- Optional Resident Check-in
- Back-lit pushbuttons for easy activation
- Cancel button
- Continuous supervision of station
- Plug-in connectors for easy service
- Optional antimicrobial Bio-seal to cover buttons and allow for easy disinfecting of the station

SPECIFICATIONS

Power Required: 7mA @ 15VDC

Power Required when used with SLIM Speaker Module: 10mA @ 15VDC

Weight: 0.22 lbs (0.09 kg)

Size: W: 3.2" (8.1 cm)

H: 4.25" (10.8 cm)

D: 0.6" (1.5 cm) in back box: 0.5" (1.3 cm) off-wall

Housing and Finish: High-impact molded Cycloy 2950 UL 94-5VB; light grey faceplate

Controls/Indicators: Two (2) Push buttons for call actuation, One (1) Call-assurance LED, One (1) Call-assurance /Monitor LED, and One (1) Cancel button

Terminations: Plug-on connectors

Backbox Requirement: 1-gang, Non-gangable W: 2.25" (5.7 cm) H: 4" (10.2 cm) D: 3.5" (8.9 cm)

Backbox Requirement when used with SLIM Speaker Module: 3-gang, Non-gangable W: 5.5" (14 cm) H: 3.75" (9.5 cm) D: 3.5" (8.9 cm)

Certification: UL/C-UL 1069, FCC Part 15 Class B, IEC60601-1-2, Designed for compliance to IEC60601-1 and CE Marking (MDD), FDA Registered, and RoHS Compliant

DESCRIPTION

The Responder® 4000 SLIM Push for Help/Staff Registration or Resident Check-in Station is an emergency dual pushbutton station that has the ability to place two (2) unique call priorities from a single station. The bottom button is used to place a "Staff Presence" or "Staff Registration" call. Using an optional button, the station can be customized to provide a "Resident Check-in" call in priority level. The Cancel button allows for cancellation of the call at the station. This type of station is normally located at a patient room head wall for placing a call for help, staff assist, and/or staff registration. When

customized with the "Resident Check-in" button, the station may be located in the bathroom or near the entrance to the unit.

The Push for Help / Staff Registration or Check-in Station is connected to its associated corridor light and is continually supervised for station presence. Through programming, a "Trouble" message can annunciate at the associated console(s) when the station is removed from the system.

*Architect and Engineer (A&E) Specifications available online at: customerconnection.rauland.com
 Specifications subject to change without notice*

©Copyright 2009 Rauland-Borg Corporation Printed in USA Rev 03/2009